

Standardization of autoimmune tests: successes and challenges

The International Federation of Clinical Chemistry and Laboratory Medicine (IFCC)
Harmonisation of Autoantibody Testing Working Group

Dr. Joanna Sheldon
Protein Reference Unit
St. George's Hospital
London




With thanks to

- The patients who generously donated their samples
- IFCC and IRMM
- The members of the WG-HAT
 - Ingrid Zegers (IRMM)
 - Allan Wiik
 - Pier Luigi Meroni
 - The companies for their support and participation
- Dr. Heinz Schimmel, Dr. Evanthia Monogioudi, Dr. Gustavo Martos-Sevilla Dr. Dana Hutu from the IRMM
- Dr. Emma Tuddenham from St. George's

Autoimmune testing..... what are we trying to do?


- ★ detect or quantify
 - ★ IgG antibodies (or IgA, IgM)
 - ★ to cell or tissue components
“antigens”
- ★ support or exclude diagnosis
 - ★ monitor disease
 - ★ suggest prognosis



Detect..... or quantify




Various substrates

- ethanol fixed neutrophils
- HEp2 cells
- Monkey kidney




Reported as

- Neg/pos
- Pattern
 - Homogeneous, speckled
 - c-ANCA or p-ANCA
- Titre or weak, strong, very strong etc.
- Subjective
- Skilled
- Hard to automate



Follow-up testing

- (more) specific
- ELISA based assays



Various substrates

- Purified
- Recombinant

Various methods

- ELISA based assays
- Multiplex assays

Reported as

- Number (concentration) with a Ref. range

Advantages and disadvantages


- Less subjective
- Easier to automate
- No standardisation
- Arbitrary values (units IU/ml, IU/L, U/ml, U/L)
- Values infer information that is not supportable
 - Patients with the same “concentration” of antibody may have completely different clinical features
 - Higher concentration worse disease is not true for many auto- antibodies
- Various ref. ranges and clinical “cut-off” values
- Marked methodological variation

Is there a problem with quantification?

Used with permission of UKNEQAS

Antibodies to myeloperoxidase, known positive sample
– distribution of method means (n=38)

- ★ Patients and clinicians move from one hospital to another
- ★ A positive results potentially varying by 10x or 100x or 1000x is NOT SAFE
- ★ clinicians may not know of this variability
- ★ patients get different results depending on where the samples is analysed



Range of method means for IgG anti MPO concentrations U/ml or IU/ml

Autoantibody testing.... the challenges

We use arbitrary units because then all our assays look the same

Does it really matter...its OK, we understand the results

Antibody – variations between patients, during disease, affinity and avidity, comparability with assay standard etc.

Antigen variation
- purified, synthetic, degraded, lot to lot variation

No robust reference materials

Detection system
- IgG, IgG & IgM, IgA, IgG subclasses, reactivity of detection antibody

Method variation
- dilution, diluent, manual, automated, conjugate, capture, direct etc.

*It will never work
We have bigger worries with glucose or TSH or...*

*It is too complicated
...and we need to use this method/analyser*

Challenge 1 – antibody

Binding of antibodies to antigens is variable – affinity and avidity

- ★ some patients make high affinity antibodies that bind very tightly
 - form stable complexes in vitro and in vivo
 - often are damaging e.g. through complement activation
 - are resilient to changes in temperature, ionic strength, pH etc.
- ★ some patients make low affinity antibodies that do not bind tightly
 - do not form very stable complexes
 - not so damaging
 - the complex can be separated by minor changes in temperature, ionic strength, pH etc.
- ★ the behaviour is not consistent through the disease course
- ★ the antibody used to “standardise” the method is unlikely to be representative of all patients auto-antibodies
- ★ QC materials are unlikely to be representative of patients samples

Challenge 2 – antigen

★ Purified

- extracted from mammalian tissue
- purification with heat, cold, salt, alcohol etc. may alter structure or denature
- contaminated with other proteins and antigens
- stability of preparations
- reproducibility of preparations
- expression of relevant antigenic epitopes


★ Synthetic

- not necessarily identical to native (structurally or antigenically)
- may lack important epitopes

★ Variability

- Between manufacturers
- Between lots

Challenge 2 – antigen e.g. proteinase 3



- ★ 3 important epitopes
- ★ At diagnosis, patients showed antibody reactivity to multiple parts of the molecule
- ★ In remission, reactivity diminished
- ★ During relapse, antibody reactivity changed e.g. from c-term to n-term
- ★ Some patients showed multiple changes throughout their disease course
- ★ Orientation of the proteinase 3 antigen in the assay is important

Clinical and Experimental Immunology ORIGINAL ARTICLE doi:10.1111/1365-2214.12009.02063.x

Epitope shift of proteinase-3 anti-neutrophil cytoplasmic antibodies in patients with small vessel vasculitis

D. Selga, M. Segelmark, L. Gunnarsson and T. Hellmark
 Department of Nephrology, Clinical Sciences in Lund, Lund University and Lund University Hospital, Sweden

the Journal of Immunology

Mapping of Conformational Epitopes on Human Proteinase 3, the Autoantigen of Wegener's Granulomatosis

Angelika Kuhl, Brice Korkmaz, Bert Utecht, Andrea Kneupert, Ulf Schönemarch, Ulrich Specks and Dieter E. Jenne

This information is current as of March 14, 2016.

J Immunol 2010; 185:387-399. Prepublished online 7 June 2010; doi: 10.4049/jimmunol.0903887
<http://www.jimmunol.org/content/185/1/387>

Challenge 3 – method variation

Immunoassay

★ ~40 different methods for IgG anti proteinase 3 in UKNEQAS (including “in house”, “others” and “not stated”)

- ★ Manual ELISA
- ★ Automated ELISA
- ★ Automated variants of ELISA
- ★ Multiplex analysis

Various

- ★ sample dilution
- ★ Diluent – e.g. variations in ionic strength
- ★ “capture” – antigen specifically bound to “well” to increase sensitivity
- ★ “capture” – antigen specifically bound and orientated on the well to expose important epitopes and increase specificity and sensitivity
- ★ direct ELISAs
- ★ Combination of rapid (minutes) and slow (hours) methods

Challenge 4 – detection system

★ What is detected?

★ IgG

★ IgG and IgM

★ IgA

Possible variation in reactivity between

★ Classes of Ig

★ Subclasses of IgG

★ between standards and patient samples reacting to the detection antibody

Robust reference material for the IgG antibody to the antigen

Where to start?
Likely to be more
than 1 step

Antigen – may need
more detailed
characterisation or
definition

Detection system

Method – may need
more detailed
characterisation or
definition

IFCC/IRMM

Harmonisation of Autoantibody Testing Working Group WG-HAT

- ★ Formed in 2010
- ★ A joint project between the IFCC and IRMM
- ★ Bring the excellence of the IRMM in preparation, analysis and validation of reference materials to autoimmune serology testing
- ★ Use similar rigorous protocols as were used on the preparation of ERM DA 470k (protein ref material)

IFCC/IRMM WG-HAT


Identified 5 analytes where the CONCENTRATION was likely to be important – IgG anti:

- ★ Myeloperoxidase
- ★ Proteinase 3
- ★ Glomerular basement membrane
- ★ Cyclic citrullinated peptide
- ★ Cardiolipin/B2 GP1 antibodies

Define the protocol for future use

What do we expect of a lab test?

Precise, Accurate, Timely, Clinically useful, CORRECT



Easy analytes e.g. glucose, calcium, where there analyte is well defined and simple

Where we want to be for Autoimmune Serology

Where we are for Autoimmune serology

Difficult analytes e.g. proteins where defining the exact composition is complicated

IgG anti MPO The process - briefly

- ★ The raw material: a plasmapheresis material from a patient with antibodies to myeloperoxidase (and relevant clinical findings)
- ★ Plasma converted into serum by the addition of protamine sulphate solution, incubation and centrifugation to remove the fibrin
- ★ Delipidation by incubation with synthetic amorphous silica
- ★ Dialysis against isotonic saline
- ★ pH adjustment
- ★ Preservatives added (sodium azide, benzamidine hydrochloride monohydrate and aprotinin)
- ★ Sterilised through a 0.22µm filter
- ★ 1ml serum transferred into vials under clean room conditions and lyophilised



- ★ Evaluation process

Characteristics of a Reference Material and ERM DA 476/IFCC

Characteristic	Explanation	ERM DA 476/IFCC
Homogeneous	Low and stated variability in concentration of the measurand between vials of the material	The uncertainty contribution for potential inhomogeneity is 0.85%
Stable	The material must be stable over its expected life-span	The material is stable e.g. during shipment (up to 2 weeks) and the on storage at -20oC and -70oC
Traceable	Related to a higher order reference material (usually national or international) through an unbroken chain of comparisons, all with stated uncertainty	
Commutable	The characteristic of a reference material to behave in a comparable way to the samples (relevant to the intended use of the reference material)	
Safe	Chemically and biologically safe (including tested as negative for HIV and Hepatitis B).	The raw material was tested and confirmed as negative for HIV, Hepatitis B and C
Ethical	Where relevant, samples from patients have been collected ethically and with appropriate agreement from the patients.	Consent given by patients for their material to be used
Available	There must be sufficient material that is readily available to relevant laboratories or companies over a time period of approx. 5-10 years. Produced with sufficient documentation to reproduce a comparable material when necessary.	Available from the IRMM
Certified	Ideally, reference material should be certified with stated uncertainties of the various characteristics	Certified in April 2015

Certified

Ideally, reference material should be certified with stated uncertainties of the various characteristics

JOINT RESEARCH CENTRE
Institute for Reference Materials and Measurements

CERTIFICATE OF ANALYSIS

ERM[®] - DA476/IFCC

HUMAN SERUM		
Mass Concentration		
	Certified value ²⁾ [mg/L]	Uncertainty ³⁾ [mg/L]
anti-MPO IgG ¹⁾	84	9

1) Anti-myeloperoxidase immunoglobulin G as measured by immunoassays


2) Unweighted mean value of the means of accepted data sets, each set obtained in a different laboratory and/or with a different method of determination. The certified mass concentration and its uncertainty are traceable to the stated value of the mass concentration in United States National Reference Preparation (USNRP) 12-08750 (Reimer et al., Am. J. Clin. Pathol. 77 (1982) 12-19).

3) The uncertainty of the certified value is the expanded uncertainty with a coverage factor $k = 2$ corresponding to a level of confidence of about 95 % estimated in accordance with ISO/IEC Guide 98-3, Guide to the Expression of Uncertainty in Measurement (GUM:1995), ISO, 2008


This certificate is valid for one year after purchase.
Sales date:
The minimum amount of sample to be used is 10 µL.

NOTE
European Reference Material ERM[®]-DA476/IFCC was produced and certified under the responsibility of the Institute for Reference Materials and Measurements of the European Commission's Joint Research Centre according to the principles laid down in the technical guidelines of the European Reference Materials[®] co-operation agreement between BAM-IRMM-LGC. Information on these guidelines is available on the internet (<http://www.erm-cm.org>).

Accepted as an ERM[®], Geel, February 2015
Latest revision: April 2015

Signed: 

Prof. Dr. Hendrik Emons
European Commission
Joint Research Centre
Institute for Reference Materials and Measurements
Retiesweg 111
B-2440 Geel, Belgium

 Registration No. 200-01M
ISO Guide 35 for the
production of reference materials

All following pages are an integral part of this certificate.
Page 1 of 3

ERM-DA476/IFCC

- ★ IgG anti MPO
- ★ Certified value 84mg/L
- ★ Uncertainty 9mg/L

IgG anti MPO Traceable

Related to a higher order reference material (usually national or international) through an unbroken chain of comparisons, all with stated uncertainty

The International Unit – only usable with WHO support

- ★ used to compare the biological activity of different preparations of the same basic substance e.g. vitamins, hormones, vaccines etc.
- ★ The mass or volume that constitutes one International Unit **varies** based on which substance is being measured
- ★ The WHO Expert Committee on Biological Standardisation provides a reference preparation of the agent, **arbitrarily** sets the number of IUs contained in that preparation, and specifies a biological procedure to compare other preparations of the same agent to the reference preparation.
- ★ The number of IUs contained in a new substance is **arbitrarily** set, there is no equivalence between IU measurements of different biological agents
 - ★ Vitamin A: 1 IU is the equivalent of 0.3 µg retinol, or 0.6 µg beta-carotene Vitamin C: 1 IU is 50 µg L-ascorbic acid
- ★ Does the “arbitrary” International Unit meet our need for a TRACEABLE reference material? Is there anything that can?

ERM-DA470k/IFCC

- ★ Produced by the IRMM
 - ★ Collaboration with Dade Behring (Marburg) and 20 laboratories across Europe
- ★ ERM-DA470K/IFCC distributed under strict transport guidelines to participating labs
- ★ Value transfer protocol detailed and strict
 - ★ *Storage, reconstitution, pipettes, balances, volumes, timing, operators, reagents, QC, assay performance etc.*
- ★ Closed and open systems used for value transfer
- ★ Specific investigations on particular issues



CERTIFICATE OF ANALYSIS

ERM[®] - DA470k/IFCC

Proteins in the reconstituted material ¹⁾	Mass concentration	
	Certified value ²⁾ [g/L]	Uncertainty ³⁾ [g/L]
α_2 macroglobulin (A2M)	1.43 ⁴⁾	0.06
α_1 acid glycoprotein (AAG)	0.617 ⁵⁾	0.013
α_1 antitrypsin (AAT)	1.12 ⁵⁾	0.03
albumin (ALB)	37.2 ⁴⁾	1.2
complement 3c (C3c)	1.00 ⁴⁾	0.04
complement 4 (C4)	0.162 ⁴⁾	0.007
haptoglobin (HPT)	0.889 ⁴⁾	0.021
immunoglobulin A (IgA)	1.00 ⁴⁾	0.05
immunoglobulin G (IgG)	9.17 ⁴⁾	0.18
immunoglobulin M (IgM)	0.723 ⁴⁾	0.027
transferrin (TRF)	2.36 ⁵⁾	0.08
transferrin (TTR)	0.220 ⁵⁾	0.018

¹⁾ When the material is reconstituted according to the specified procedure (see page 3).
²⁾ The certified values are the unweighted means of 6-14 accepted mean values, independently obtained by 5-14 laboratories, using ERM-DA470 as calibrant (Baudner et al., EUR reports 15423 and 16882 European Communities, Luxembourg (1993)).
³⁾ Expanded uncertainty with a coverage factor $k = 2$ corresponding to a level of confidence of about 95 % estimated in accordance with the Guide to the Expression of Uncertainty in Measurement (GUM), ISO, 1995.
⁴⁾ This certified mass concentration is traceable to the stated value of the mass concentration in USNRP 12-0575C (Reimer et al., Am. J. Clin. Pathol. 77 (1982) 12-18) used as calibrant for assigning values to ERM-DA470, applying the procedures described for the certification of ERM-DA470 and in the report for ERM-DA470k/IFCC.
⁵⁾ The certified value in the calibrant ERM-DA470 was obtained by calibration with a pure protein preparation (Strup-Jensen, Clin. Chem. Lab. Med. 39 (2001) 1099 - 1097). Consequently, the certified value in ERM-DA470k/IFCC is traceable to the International System of Units (SI) via ERM-DA470, applying the procedures described in the certification report of ERM-DA470 (see point 2) and in the report for ERM-DA470k/IFCC.

This certificate is valid for one year after purchase.

Sales date:

The minimum amount of sample to be used is 2 μ L.

Accepted as an ERM[®], Geel, July 2008.

Signed:

Prof. Dr. Hendrik Emons
Unit for Reference Materials
EC-DG JRC-IRMM
Rortseweg 111
2440 Geel, Belgium



Registration No. 200-TEST
ISO Guide 34 for the
production of reference materials.

All following pages are an integral part of the certificate.

Page 1 of 4

IgG anti MPO Traceable

Related to a higher order reference material (usually national or international) through an unbroken chain of comparisons, all with stated uncertainty

- ★ We are measuring IgG.....with specific antibody activity against myeloperoxidase

- ★ The value assignment of IgG anti MPO was done using:
 - ★ with dilutions of the candidate reference materials
 - ★ Purified IgG anti MPO
 - ★ affinity chromatography using a protein A column
 - ★ Hi-trap column using purified human myeloperoxidase
 - ★ Superdex 200 10/300 column
 - ★ Confirmation of purity of material
 - ★ Dilutions of ERM-DA470k/IFCC (CRM for IgG)


- ★ These materials were measured under strict protocols by a variety of methods

IgG anti MPO Value assignment


- ★ The affinity purified Abs or monoclonals can be assigned values that are traceable to the SI (via traceability to ERM-DA470k or UV-absorption measurements) - **VITAL**
- ★ They can be used to make the values in the matrix material traceable to the SI.
- ★ Certified values 84 mg/L (uncertainty 9mg/L)

IgG anti MPO Commutable

The characteristic of a reference material to behave in a comparable way to the samples (relevant to the intended use of the reference material)



Preliminary commutability study for Myeloperoxidase antibodies



Numerical recalibration of values for clinical samples using a conversion factor based on results for a candidate reference material (RM 5)

- good convergence for 6 out of 7 methods
- outliers remain and become more evident
 - this problem can not be solved by recalibration

IgG anti MPO Commutable

The characteristic of a reference material to behave in a comparable way to the samples (relevant to the intended use of the reference material)

- ★ Different formats of the reference material, all based on the same raw material have been tested and have been shown to be commutable for combinations of SEVEN methods
- ★ It is expected that ERM-DA476/IFCC will be commutable for the majority of IgG anti MPO methods
- ★ If another method is used, then commutability should be verified

IgG anti MPO Commutable

The characteristic of a reference material to behave in a comparable way to the samples (relevant to the intended use of the reference material)



Correlation coefficients 2nd commutability study

	Wieslab C	Phadia EliA	Euro-immune	Varelisa	Orgentec	Quanta Lite	IMMCO	Biorad EIA	Bioflash	Aesku
Bioplex	0.29	0.71	0.58	0.60	0.55	0.79	0.69	0.65	0.71	0.39
Wieslab C		0.69	0.74	0.84	0.74	0.77	0.55	0.65	0.59	0.79
Phadia EliA			0.83	0.88	0.90	0.94	0.79	0.95	0.90	0.84
Euroimmune				0.79	0.90	0.84	0.66	0.79	0.65	0.80
Varelisa					0.83	0.95	0.79	0.90	0.85	0.87
Orgentec						0.88	0.74	0.89	0.79	0.95
Quanta Lite							0.85	0.93	0.87	0.89
IMMCO								0.88	0.81	0.72
Biorad EIA									0.92	0.84
Bioflash										0.74

Clinical interpretation of results using cut-offs provided by manufacturers

averages	W Capture	IMMCO	BioRad EIA	Orgentec	EliA_Phadia	QuantaLite	Inova BIOFLASH	AESKULISA	Varelisa	Euroimmun	Bioplex2200
1137.77			2540.66		331.38		739.80	177.96			1899.06
133.19	108.23	39.97	191.23	41.02	80.02	57.96	406.91	145.72		168.76	92.07
110.91		31.08	147.05	47.82	70.56	61.88	187.29	259.49		185.37	7.70
109.91	66.11	47.53	126.41	23.30	68.34	55.42	384.00	142.57		117.06	68.32
98.02	42.87	32.78	93.00	35.70	66.67	40.89	234.95	113.55	82.27	183.06	152.48
97.91	88.13	21.93	87.14	47.43	44.90	40.04	252.38	238.29	80.98	160.64	15.14
81.67	58.71	33.50	90.53	17.22	39.85	48.20	295.07	120.78	88.27	71.87	34.36
73.86	157.21	21.28	57.03	22.92	33.82	40.73	101.81	126.87	93.20	149.96	7.61
70.76	124.07	17.78	60.49	25.17	49.78	43.64	126.47	123.39		132.71	4.11
37.47	44.04	14.98	25.73	11.48	12.96	21.59	138.91	72.58	27.47	38.25	4.21
32.30	53.45	8.27	27.03	10.37	7.63	17.33	59.42	63.19	31.57	73.23	3.78
31.32	50.95	13.32	18.43	8.97	7.99	22.06	25.60	71.47	44.82	77.24	3.64
30.11	16.94	12.48	22.57	8.32	12.78	19.35	95.13	51.00	21.70	45.73	25.24
28.85	18.08	7.83	16.03	11.17	7.89	19.78	49.84	60.20	24.02	94.66	7.84
28.33	12.52	8.50	47.79	13.28	8.97	17.34	44.28		23.22	99.19	8.23
28.18	12.24	22.00	35.91	12.25	19.44	11.74	83.36	79.33	16.55	13.51	3.62
23.51	5.63	10.57	18.19	6.02	17.50	11.90	67.12	28.32	17.92	67.06	8.37
22.21	36.30	11.17	11.22	6.52	5.55	15.46	16.26	40.62	35.07	64.38	1.77
21.85	59.30	6.00	8.76	6.80	10.33	14.73	8.84		33.82	69.36	0.57
17.58	7.40	9.98	7.01	5.42	9.17	19.25	14.70	26.70	16.47	73.45	3.86
14.58	7.37	5.90	5.95	5.10	4.21	12.91	21.66	8.31	17.33	66.47	5.13
12.14	5.09	5.60	13.31	3.07	3.45	11.51	32.36		19.85	25.30	1.87
11.76	5.09	14.23	18.50	3.80	14.77	18.60	3.36		4.12	33.39	1.70
10.98	9.78	4.23	5.99	2.70	5.69	11.29	28.84		10.37	27.20	3.68
9.22	7.10	3.88	1.31	2.47	2.77	10.32	7.99	6.77	5.50	51.70	1.60
7.58	2.44	5.33	1.07	1.98	1.34	9.39	4.88	7.67	8.70	39.40	1.13
7.51	2.73	5.63	2.02	1.28	1.26	5.26	11.82	6.96	8.65	35.86	1.13
4.52	3.23	5.02	1.24	1.22	3.07	5.04	11.18	1.37	4.47	10.89	3.04
3.92	0.38	20.82	5.16	1.20	0.15	3.33	3.20	3.61	1.32		0.02
2.72	0.63	5.78	6.00	1.02	2.03	3.27	3.20	4.03	1.17		0.05

9 samples with the same interpretation in all methods, 13 in 10 out of 11 methods

Table 1 | Comparison of methods for testing for PR3-ANCA and MPO-ANCA in ANCA-associated vasculitis

Patient population (n) vs comparison group (n)	Method	Sensitivity (%)	Specificity (%)	PPV (%)	NPV (%)	AUC/ROC	Comments
GPA (86) vs non-vasculitic disease (450) ²⁰	IFT	92	99	Nd	Nd	0.96 (0.94–0.98)	Histological diagnosis Retrospective study
	Direct PR3-ANCA ELISA	60	99			0.80 (0.76–0.83)	
	Capture ELISA	72	99.3			0.86 (0.82–0.89)	
	Anchor ELISA	96	98.5			0.96 (0.94–0.98)	
GPA (232) vs inflammatory diseases (661) ²³	IFT	77.9	90.9	73	93	Nd	Histological diagnosis Prospective study
	Anchor ELISA	80.4	97.4	88	93		
GPA (59*) vs inflammatory and infectious diseases (585) ²³	Hn-hr PR3-ANCA ELISA	94	99 (predefined)	Nd	Nd	Nd	Histological diagnosis Retrospective study
	Capture ELISA	66					
	Direct (hn) PR3-ANCA	64					
GPA (34) vs SLE (65) ²¹	Direct PR3-ANCA Anchor ELISA	97.1	98.4	Nd	Nd	0.999 (0.947–1.00)	Clinical diagnosis Retrospective study
GPA (40) vs RA or SLE (20) ²²	IFT	62.5	95–100	Nd	Nd	Nd	Histological diagnosis Retrospective study
	Direct PR3-ANCA (n=5 kits)	45–55					
	Capture ELISA (n=2 kits)	60–62.5					
	Anchor ELISA (n=4 kits)	60–62.5					
MPA (40) vs RA or SLE (20) ²²	IFT	82.5	95–100	Nd	Nd	Nd	Histological diagnosis Retrospective study
	Direct MPO-ANCA ELISA (n=8 kits)	62.5–85					
	Capture ELISA (n=2 kits)	80					
	Anchor ELISA (n=1 kit)	75					
GPA (55) vs suspected vasculitis (175) ²²	IFT	69.1	100	Nd	Nd	Nd	Clinical diagnosis Retrospective study
	Direct PR3-ANCA ELISA (n=2 kits)	61.8–72.7	95.4–96.4			0.856–0.879	
	Capture ELISA (n=2 kits)	70.9–72.7	95.9–99.5			0.862–0.878	
	Anchor ELISA (n=3 kits)	61.8–72.7	98.5–99.0			0.833–0.881	
	Other assays (n=2)	72.7–74.5	95.9–97.9			0.878–0.902	

*47 of 59 patients in the GPA group had a cytoplasmic ANCA pattern on IFT. Abbreviations: ANCA, antineutrophil cytoplasmic antibody; AUC, area under the curve; GPA, granulomatosis with polyangiitis; hn, human native; hr, human recombinant; IFT, indirect immunofluorescence technique; MPA, microscopic polyangiitis; MPO, myeloperoxidase; Nd, not determined; NPV, negative predictive value; PPV, positive predictive value; PR3, proteinase 3; RA, rheumatoid arthritis; ROC, receiver operating characteristics; SLE, systemic lupus erythematosus.

We can improve the numbers....

- ★ Introduction and adoption of traceable commutable reference materials should reduce the variability in the values for autoantibody measurements
- ★ It will not solve the inherent variability in the values given by certain patient samples in different methods
- ★ It should help identify methodological outliers and guide investigation and improvements

Standardization in autoimmune testing

IFCC/JRC-IRMM WG-HAT

Successes

- ★ huge advances
- ★ defined processes for making CRM for autoantibodies
- ★ further materials in progress

Challenges

- ★ introducing the materials
- ★ evaluate the impact e.g. on patient and EQA
- ★ consider further harmonisation or better definition of:
 - ★ antigen type/source, method, detection system